

NIST No. 1848 (Lubricating Oil Additive Package)

Sheet No.: **NSX2100V-CH-008E** Petroleum chemistry
 Measurement model: **TS-2100V System** Relevant standard: ASTM D5453
 ASC-250L/VF-210/SD-210 JIS K 2541
 Detection method: **Ultraviolet Fluorescence method**

As the sulfur in lubrication oil is a direct cause of corrosion at metallic part of machine, it needs to be controlled at low concentration. The sulfur analysis device (**TS-2100V**) of Mitsubishi Chemical Analytech Co., Ltd. can analyze the sulfur in lubrication oil quickly with accuracy.

Sample name	Lubricating Oil Additive Package																						
Analytical item	Quantitative analysis of sulfur in combustion method																						
Standard	ASTM-D5453 : standard testing method for measuring sulfur contained in carbon hydride and fuel using an ultraviolet fluorescence detector JIS K 2541 : raw petroleum and petroleum product – sulfur content testing method – ultraviolet fluorescence method																						
Analytical principle	Ultraviolet fluorescence method: Sample is burned in argon / oxygen stream and the generated sulfur dioxide is introduced to the cell of ultraviolet irradiation. The fluorescence intensity generated by ultraviolet irradiation is measured and the amount of sulfur is calculated based on the standard curve that has been created using the standard sulfur sample. $\text{Organic-S} + \text{O}_2 \rightarrow \text{SO}_2 + \text{CO}_2$ (combustion) $\text{SO}_2 + \text{h}\nu \rightarrow \text{SO}_2 + \text{h}\nu_2$ (ultraviolet fluorescence)																						
Result of sulfur analysis	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th rowspan="2">Sample name</th> <th colspan="5">TS-2100V analysis value (S %)</th> </tr> <tr> <th>1</th> <th>2</th> <th>3</th> <th>Average</th> <th>RSD (%)</th> </tr> </thead> <tbody> <tr> <td>Lubricating Oil Additive Package</td> <td>2.278</td> <td>2.305</td> <td>2.278</td> <td>2.287</td> <td>0.67</td> </tr> </tbody> </table>						Sample name	TS-2100V analysis value (S %)					1	2	3	Average	RSD (%)	Lubricating Oil Additive Package	2.278	2.305	2.278	2.287	0.67
Sample name	TS-2100V analysis value (S %)																						
	1	2	3	Average	RSD (%)																		
Lubricating Oil Additive Package	2.278	2.305	2.278	2.287	0.67																		
Vertical combustion method																							
Required analysis time	Pretreatment (---) minutes, Measurement (4) minutes																						
Vertical type	Total (4) minutes/ (1)measurement																						

*This sheet is provided as a reference and does not guarantee analytical values. Optimal conditions may vary depending on external factors, such as the analysis environment, and the nature of the sample.

<p>Conditions of sulfur analysis</p> <p>Vertical combustion method</p>	<p>Measurement condition</p> <p>Temperature of electric furnace</p> <p style="padding-left: 40px;">Inlet Temp 900°C</p> <p style="padding-left: 40px;">Outlet Temp 1,000°C</p> <p style="padding-left: 40px;">PMT Range Low (for High concentration)</p> <p>Gas flow rate</p> <p style="padding-left: 40px;">Ar..... 100mL/min</p> <p style="padding-left: 40px;">O₂ 500mL/min</p> <p style="padding-left: 40px;">Ar time: 30sec</p> <p style="padding-left: 40px;">O₂ time: 120sec</p> <p>Standard sample for standard curve: S_Dibutyl disulfide / toluene 0, 1, 10, 50, 100µg/mL × 30µL</p> <p>Amount of introduced sample: 20µL</p> <p>- For the sample for measurement, the sample diluted to 0.25% with toluene was used. The obtained measurement value multiplied by dilution rate was set to the sulfur quantitative value.</p>
--	--